

Limulus Amebocyte Lysate

PYROTELL®-T

Manufactured by: Telephone: (508) 540-3444
ASSOCIATES OF Telephone: (888) 395-2221
CAPE COD Toll-Free: (888) 395-2221
INCORPORATED Fax: (508) 540-8680
124 Bernard E. Sant Jun Drive • E. Falmouth, MA 02536 USA
Technical Support: (800) 848-3248
Customer Service: (800) 525-8378

PN000845 rev6 April 2017

For The Detection And Quantification Of Gram Negative Bacterial Endotoxins (Lipopolysaccharides)

Pyrotell®-T *Limulus* amebocyte lysate (LAL) reagent is intended for the *in vitro* detection and quantitation of endotoxin in the end-product testing of human injectable drugs (including biological products), animal injectable drugs, and medical devices. It may also be used for testing raw materials (including water) and components used in production and for in-process monitoring of endotoxin levels. The USP Bacterial Endotoxins Test (1) is the official test referenced in specific USP monographs. Use Pyrotell®-T for *in vitro* diagnostic purposes only. Pyrotell®-T is not intended for detection of endotoxin in clinical samples or for diagnosis of disease in humans or animals.

Summary of Test

Limulus amebocyte lysate (LAL) is an aqueous extract of blood cells (amebocytes) from the horseshoe crab, *Limulus polyphemus*. In the presence of endotoxin LAL becomes turbid and, under appropriate conditions, forms a solid gel-clot. The turbidimetric LAL test is performed by adding a given volume of Pyrotell®-T to a given volume of specimen and incubating the reaction mixture at 37°C. The higher the endotoxin concentration in the specimen, the faster turbidity will develop. Turbidity can be used to quantify endotoxin concentration in two ways.

In the kinetic turbidimetric LAL method, either the rate of increase in turbidity or the time taken to reach a particular level of turbidity (the onset time) is determined. Higher endotoxin concentrations give shorter onset times. The assay requires specialized instrumentation to incubate multiple samples at a controlled temperature (usually 37°C) and take optical density readings at regular intervals. Standard curves may be constructed by plotting the log onset time against the log concentration of standard endotoxin and are used to calculate endotoxin concentrations in specimens.

The turbidimetric LAL method is rapid, specific, easy to perform, and highly sensitive. The detection limit depends on the method and instrumentation used and may be as low as 0.001 Endotoxin Units (EU) per mL.

History and Biologic Principle

Howell described the clotting of *Limulus* blood in 1885 (2). In the 1950's, Bang at the Marine Biological Laboratory, Woods Hole, MA, discovered that gram negative bacteria cause *Limulus* blood to clot (3). Levin and Bang later determined that the reaction is enzymatic and that the enzymes are located in granules in the amebocytes (4). They showed that clotting is initiated by a unique structural component of the bacterial cell wall called endotoxin or lipopolysaccharide (5). Current

understanding is that the reaction consists of a cascade of enzyme activation steps. While the complete reaction is not understood, the last step is well described. Clotting protein (coagulogen) is cleaved by activated clotting enzyme; the insoluble cleavage products coalesce by ionic interaction and the turbidity of the reaction mixture increases. The greater the amount of endotoxin present, the faster turbidity develops. More information about the LAL reaction and applications is available in the literature (6, 7).

Reagent

Pyrotell®-T is packaged in lyophilized form. Pyrotell®-T contains an aqueous extract of amebocytes of *L. polyphemus*, human serum albumin (stabilizer), NaCl and other appropriate ions. No preservatives or buffers are added.

Pyrotell®-T is not labeled with a specific sensitivity. Sensitivity in a given test (designated λ) is the lowest endotoxin concentration used to construct the standard curve. In the Pyros Kinetix Flex® (Associates of Cape Cod, Inc.) the detection limit, and thus the lowest possible value of λ , is 0.001 EU/mL.

The toxicity of this reagent has not been determined; thus, caution should be exercised when handling Pyrotell®-T.

Reconstitute Pyrotell®-T as follows:

1. Gently tap the vial of Pyrotell®-T to cause loose LAL to fall to the bottom. Break the vacuum by lifting the gray stopper. Do not contaminate the mouth of the vial. Remove and discard the stopper; do not inject through or reuse the stopper. A small amount of LAL on the stopper will not affect the test.
2. Reconstitute Pyrotell®-T with LAL Reagent Water (LRW, see below) or compatible buffer (available from Associates of Cape Cod, Inc.). Add 5 mL as indicated on the vial label. Once reagent is dissolved, gently swirl the vial to insure homogeneity. Mixing too vigorously may cause excessive foaming which can cause a loss of sensitivity. Cover the vial with Parafilm M® (Bemis Company Inc.™) when not in use.

Storage Conditions

Freeze-dried Pyrotell®-T is relatively heat stable and, if kept refrigerated, will retain full activity through the expiration date on the label. Store the product at -20 to +8°C. Temperatures below -20°C can shrink the stopper, leading to a possible loss of vacuum and contamination of Pyrotell®-T. Temperatures in excess of 37°C can cause rapid deterioration of lyophilized Pyrotell®-T as evidenced by loss of sensitivity and a distinct yellowing of the product. Pyrotell®-T is shipped with cold packs in insulated containers to protect against high temperatures.

Reconstituted Pyrotell®-T is usually clear and slightly opalescent. An occasional lot will exhibit a slight, uniform turbidity. The presence of small fibers or strands does not indicate contamination nor affect activity; however, flocculent precipitation or a distinct yellow color does indicate deterioration.

Reconstituted Pyrotell®-T is less stable than the freeze-dried product; vials can be held for up to 24 hours at 2 to 8°C. Reconstituted Pyrotell®-T can be frozen once. It will retain activity for three months if frozen immediately after reconstitution and held at or below -20°C. After thawing, the same visual criteria for quality are applied as for initial reconstitution.

Specimen Collection and Preparation

Specimens should be collected aseptically in non-pyrogenic containers. Reused, depyrogenated glassware or sterile,

disposable, polystyrene plastics are recommended to minimize adsorption of endotoxin to container surfaces. Not all plastic containers are free of detectable endotoxin and an extractable substance from some types may interfere with the LAL test. The use of Pyrosol® Buffer will often overcome variation seen from interference from consumables. Labware can be tested for acceptability by randomly selecting containers from a batch, rinsing them with a small volume of LRW at room temperature for one hour, and testing the rinse as a specimen. The rinse should contain significantly less endotoxin than the lowest standard concentration to be used. Also, the rinse should neither inhibit nor enhance the test as determined by recovery of a known amount of added endotoxin.

The pH of the reaction mixture should be 6 to 8. Adjust the pH of the specimen with HCl, NaOH, or buffer (free of detectable endotoxin). Dilute concentrated HCl or NaOH with LRW and use in normalities that will not lead to significant dilution of the test specimen. If a precipitate forms in the sample upon pH adjustment, dilute the sample (not to exceed the MVD – see “Limitations of Procedure”) before adjusting the pH. Alternatively, reconstitute the Pyrotell®-T with a compatible buffer (available from Associates of Cape Cod, Inc.) and check the pH of the reaction mixture. Do not adjust the pH of unbuffered saline or water. Note that dilution alone may overcome pH problems.

Substances that denature proteins, chelate ions, bind endotoxin, or alter endotoxin's hydrophobic state may interfere with the test. Interference may be detected as recovery of significantly more or less endotoxin than expected when a known amount of standard endotoxin is added to the specimen (see “Limitations of Procedure”). In most cases, dilution of the specimen will reduce the concentration and activity of interfering substances and still yield valid test results. Appropriate controls and dilution schemes are discussed under “Test Procedure.”

Specimens should be tested as soon as possible after collection. It may be advisable to freeze non-sterile specimens that will be stored or shipped before testing. Specimens expected to contain low concentrations of endotoxin (less than 1 EU/mL) should be tested for loss of endotoxin during storage.

Test Procedure

Test Reagents

1. Pyrotell®-T (see description and method of reconstitution above).
2. *LAL Reagent Water* (LRW), not provided with Pyrotell®-T; order separately. Lyophilized Pyrotell®-T must be reconstituted with water (or compatible buffer available from Associates of Cape Cod, Inc., see item 3 below) that shows no detectable endotoxin in the LAL test. Recommended sources for water include Associates of Cape Cod, Inc. or any commercially available USP Sterile Water for Injection (sterile WFI) without bacteriostat or USP Water for Irrigation. Any of these may be used provided they have been shown to be acceptable for use as an LRW. The endotoxin limit for USP sterile WFI is only 0.25 EU/mL; therefore, sterile WFI may contain detectable endotoxin and be unsuitable for use.

To certify that water is acceptable as an LRW, test it as a specimen with a positive product control (see item 1.c. in the section on “Controls”). Use certified LRW to reconstitute Pyrotell®-T, to make dilutions of standards, and to prepare positive controls (see items 1.a. and 1.b. under “Controls”). Construct a standard curve from the onset times for the standards. The correlation coefficient should be at least 0.980 (absolute value). The endotoxin concentration of the water being tested can be estimated

by extrapolation of the standard curve below the lowest endotoxin concentration and should be less than that of the lowest standard. Also, the endotoxin concentration of the positive product control should be within 50% to 200% of the nominal concentration of the added endotoxin “spike.”

3. *Buffer*, not provided with Pyrotell®-T; order separately if required. Pyrosol buffer (Cat# BC051 or BC554) or Glucashield™ buffer (Cat# GB051) can be used instead of LRW to reconstitute Pyrotell®-T to help overcome a sample pH problem, consumable interference or interference from glucans.
4. *Standard Endotoxin*, not provided with Pyrotell®-T; order separately. Control Standard Endotoxin (CSE), obtained from Associates of Cape Cod, Inc., is used to construct standard curves, validate product, and prepare inhibition controls. Each vial contains a measured weight of endotoxin. USP Endotoxin Reference Standard may be obtained from the U.S. Pharmacopeial Convention, Inc. Follow manufacturer's directions for reconstitution and storage of standard endotoxins. CSE lots may show different potencies (EU/ng) when tested with different lots of Pyrotell®-T. If using CSE, endotoxin concentrations can be expressed in EU/mL if the potency of a given lot of CSE has been determined with the Pyrotell®-T lot in question (1, 2).

Materials and Equipment (not provided)

1. *Reaction vessels*. Choice of type depends on the instrumentation used to measure turbidity. Reaction tubes for the Pyros Kinetix Flex system are 8 x 75 mm, depyrogenated borosilicate glass tubes (TK100).
2. *Optical reader*. For the kinetic turbidimetric method, the use of an incubating optical reader such as the Pyros Kinetix Flex or the BioTek ELX 808 IU available from Associates of Cape Cod, Inc. is recommended.
3. *Test tube racks* to hold and/or incubate reaction tubes.
4. *Pipets*, automatic pipettors with pipet tips, or repeating pipettors with plastic syringe barrels. Disposable pipets and tips free of interfering endotoxins and glucans are recommended. Associates of Cape Cod, Inc. offers the Pyroclear® line, which comes with a Certificate of Compliance certifying that the products are free of interfering endotoxins and glucans.
5. *Vortex-type mixer*.
6. *Parafilm M®*. The side in contact with the paper backing is normally nonpyrogenic.
7. *Nonpyrogenic test tubes* with adequate capacity for making dilutions of endotoxin standard or test specimen. See “Specimen Collection and Preparation” for other containers suitable for dilutions.
8. *Hot air oven* with 250°C capacity for depyrogenation of glassware. Commonly used minimum time and temperature settings are 30 minutes at 250°C (1,8,9).

Controls

Controls are necessary to insure a valid test. Recommended procedures are detailed USP (1).

1. Endotoxin controls

- a. **Endotoxin standard series**. Prepare a fresh set of dilutions from the stock endotoxin solution for each test. Do not use previously prepared dilutions unless you have demonstrated the stability of that range of concentrations. Make dilutions in a geometric series (e.g. twofold, fourfold or tenfold dilutions) to give the

range of endotoxin concentrations required. Concentrations of 1.0, 0.5, 0.25, 0.125, 0.0625 and 0.03125 EU/mL are recommended to verify Pyrotell®-T performance. The lowest endotoxin concentration in any standard series is the detection limit of that particular test and is designated λ. To get to the range of standards required, use as few dilutions as possible with appropriate pipet volumes to maximize accuracy.

- b. **Positive control** (a single standard endotoxin concentration) should be included if the standard series (see a. above) is not prepared in the same way as the positive product controls (see c.). A concentration of 4λ is appropriate for standard curves constructed from 4, 5 or 6 twofold dilutions of standard endotoxin (e.g., 4λ = 0.125 EU/mL in the standard series given in section a. above). In situations where the increment of dilution is greater than twofold, the concentration of the positive control should equal that of a standard from the middle of the standard curve. For example, a value of 0.1 EU/mL would be appropriate for positive controls included with a standard series that is comprised of 0.001, 0.01, 0.1, and 1.0 EU/mL, concentrations. If a standard series is not included in a test, a positive control must be included to verify that it is appropriate to use the parameters of a previous standard curve to calculate endotoxin concentrations. Refer to the FDA guideline (1) and Interim Guidance (2) under Routine Testing of Drugs (or Devices) by the LAL Test for details.

- c. **Positive product controls** are inhibition/enhancement controls and consist of the specimen or dilution of specimen to which standard endotoxin is added. The concentration of added endotoxin in the test specimen should be the same as that of the positive control. See section b. above for selection of the appropriate endotoxin concentration for the positive product control. The added endotoxin is frequently referred to as a “spike.”

2. Negative controls

LRW negative controls should be included with each test.

Specimen Preparation - Determination of Test Dilution

If a test protocol has been developed previously for the type of specimen under test, make the necessary dilution for the assay and proceed as directed under “Performing the Test” below. If a protocol has not been developed for the specimen type, make a series of tenfold dilutions of the specimen. Do not exceed the Maximum Valid Dilution (MVD) of the product by more than a factor of 10. [Refer to the “Limitations of Procedure” below or USP (1) for explanation and calculation of MVD].

Prepare appropriate dilutions of all specimens with a positive product control for each one.

Performing the Test

Consistent technique is necessary to obtain satisfactory results. All controls and specimens should be tested in (at least) duplicate.

1. Prepare test instrumentation as necessary. In an automated system this usually involves entering sample descriptors and setting of sample test parameters.
2. Add the appropriate volume of sample (negative control, endotoxin standard, specimen, positive control or positive product control) to the reaction tube or microplate. For a 1:1 ratio in the Pyros Kinetix Flex use 0.1 mL; for a 1:4 ratio in the Pyros Kinetix Flex, use either a 0.2 mL or 0.4 mL volume.

3. Add Pyrotell®-T as appropriate for procedure. For methods using individual reaction tubes the timing of the reaction in each tube is critical. To each reaction tube in turn, add Pyrotell®-T, mix for approximately 2 seconds and place the tube into the incubating optical reader. The volume of Pyrotell®-T added is 0.1 mL for the 1:1 ratio method of the Pyros Kinetix Flex, and most other methods. For the Pyros Kinetix Flex 1:4 ratio, if the sample volume is 0.4 mL, add 0.1 mL LAL. If using a sample volume of 0.2 mL, a volume of 0.05 mL of LAL is appropriate. Failure to mix adequately is a common cause of unsatisfactory tests. Pyrotell®-T may be added most conveniently using a repeating pipettor. A fresh pipet or pipet tip is recommended for each entry into the Pyrotell®-T vial. Different volumes of Pyrotell®-T may be appropriate for other test protocols.

4. Once the incubation has started, do not disturb the reaction tube(s). The laboratory bench supporting the incubator/ optical reader should be free from excessive vibration.
5. Read the test. Allow the test to run until all samples have incubated for longer than the time required for the lowest standard endotoxin concentration to reach onset OD. Automated test systems will usually terminate the test after a preset period.

Results

1. Preliminary calculations.

Determine the time taken for specimens to reach a particular optical density threshold (usually 0.02 OD units in a tube reader) after any data corrections have been made. Optical density readings must be relative to an initial reading taken to be 0 OD units. The system software will do this. The time taken to reach the OD value is called the **onset time**.

2. Construct a standard curve.

Construct a standard curve by linear regression of the log onset time on the log endotoxin concentration for the standards. The equation for the regression line describes the standard curve (unless the option to use polynomial regression is taken as described in the following paragraph).

Provided that the absolute value of the correlation coefficient for the standard curve is at least 0.980 (see item 2. under Interpretation below), a polynomial (i.e. quadratic) regression line equation may be used to calculate endotoxin concentrations.

3. **Calculate the endotoxin concentrations.** Calculate endotoxin concentrations of all specimens (including standards and controls) using the line equation for a straight line:

$$Y = aX + b$$

where:

Y = log onset time

X = log endotoxin concentration

a = slope of the line

b = the Y intercept.

For the quadratic model the equation is:

$$Y = a_1X + a_2X^2 + b$$

Where a_1 and a_2 are the slopes of X and X²

These calculations are commonly performed by endotoxin test system software.

Interpretation

1. In order for a test to be valid, the endotoxin concentration of negative controls (estimated by extrapolation of the standard curve) should be less than that of the lowest standard concentration.
2. When a standard curve is included with the test, the absolute value of the correlation coefficient for the standard curve should be greater than or equal to 0.980.
3. The mean measured endotoxin concentration of positive controls must be within 25% of the nominal concentration. Thus, if the positive control is 0.125 EU/mL, the measured concentration must be between 0.093(75) and 0.156(25) EU/mL.
4. Determine the mean range of onset times for the standard curve. For example, if the onset times of two replicates of the highest standard endotoxin concentration are 1079 and 1087 seconds and those for the lowest concentration are 1954 and 1968, the mean range of onset times is 1083 to 1961. Valid endotoxin concentrations can only be calculated for specimens with a mean onset time that falls within the mean range of onset times of the standard curve.

For example, a valid result can be given for an unknown that gives onset times of 1949 and 1965 (mean = 1957), despite the fact that one of the replicates lies outside the standard range.

5. In order to demonstrate that the specimen does not significantly interfere with the LAL/endotoxin reaction, the measured endotoxin concentration of the positive product control must be within 50 to 200% of the nominal concentration of the added endotoxin ‘spike.’ Before determining whether the spike is recovered within these limits, subtract the endotoxin concentration measured in the (unspiked) specimen. For example, in order to be considered free of significant interference, the measured endotoxin concentration in a 0.125 EU/mL positive product control (after subtraction of any endotoxin in the unspiked specimen) must be within the range 0.0625 - 0.25 EU/mL (50 to 200% of 0.125 EU/mL). If the measured endotoxin concentration in the unspiked specimen is 0.028 EU/mL and that in the positive product control is 0.163 EU/mL, the endotoxin attributable to the spike is 0.163 - 0.028 = 0.135 EU/mL. This is within the range and subject to other requirements being met; the test of the sample is valid.

Limitations of Procedure

The procedure is limited by the extent of the inhibition or enhancement capacity of the specimen under test. If the procedure cannot be validated (1) at a specimen concentration that is greater than the minimum valid concentration (MVC), then the LAL test cannot be substituted for the USP Pyrogen Test. The MVC is calculated as follows:

$$MVC = \frac{\lambda \text{ (specimen dose)}}{\text{(endotoxin tolerance limit)}}$$

where λ is in EU/mL, specimen dose is in units per kg body weight, and the endotoxin tolerance limit is in EU/kg.

The maximum valid dilution (MVD) is the specimen dilution containing the MVC (1). It is the initial specimen concentration divided by the MVC.

The endotoxin tolerance limit is for parental drugs is specified in the individual monographs(1). The limit for medical devices is expressed per mL of an extraction or rinse volume obtained

as described in the USP (10) based on the limits for devices. For devices that do not contact cerebrospinal fluid, the limit is 20 EU/device; for those that do, it is 2.15 EU/device. The limit for liquid devices is the same as that for drugs.

Trypsin will cause a false positive result unless denatured by heat treatment before testing. Materials such as blood, serum, albumin, and plasma may interfere with turbidimetric assays.

Expected Values

Endotoxin in specimens can be quantified between the range of standard endotoxin concentrations used to construct the standard curve. If it is necessary to dilute the specimen to overcome any inhibition or enhancement, the least amount of endotoxin that can be detected will be increased accordingly. Materials derived from biological sources, even after biochemical purification, may still contain measurable levels of endotoxin. Water obtained by distillation, reverse osmosis, or ultrafiltration may contain less endotoxin than detectable as long as the purification process is operating correctly and the water is not contaminated after production.

Bibliography

1. Bacterial Endotoxin Test, chapter <85> United States Pharmacopeia (current version), United States Pharmacopeial Convention, Rockville, MD.
2. Howell, W. H. 1885. Observations upon the chemical composition and coagulation of the blood of *Limulus polyphemus*, *Callinectes hastatus*, and *Cucumaria* sp. Johns Hopkins University Circular 43:4-5.
3. Bang, F. B. 1953. The toxic effect of a marine bacterium on *Limulus* and the formation of blood clots. Biol. Bull. (Woods Hole, MA) 105:361-362.
4. Levin, J., and F. B. Bang. 1964. A description of cellular coagulation in the *Limulus*. Bull. Johns Hopkins Hosp. 115:337-345.
5. Levin, J., and F. B. Bang. 1964. The role of endotoxin in the extracellular coagulation of *Limulus* blood. Bull. Johns Hopkins Hosp. 115:265-274.
6. Levin, J., and F. B. Bang. 1968. Clottable protein in *Limulus*: its localization and kinetics of its coagulation by endotoxin. Thromb. Diath. Haemorrh. 19:186-197.
7. Progress in Clinical and Biological Research Vol. 231, Detection of Bacterial Endotoxins with the *Limulus* Amebocyte Lysate Test. 1987. Watson, S. W., J. Levin, and T. J. Novitsky (eds), Alan R. Liss, Inc., NY.
8. Tsuji, K. and S. J. Harrison. 1978. Dry heat destruction of lipopolysaccharide: Dry-heat destruction kinetics. Appl. Environ. Microbiol. 36:710-714.
9. Sweet, B. H. and J. F. Huxsoll. Depyrogenation by dry heat, ch. 12, p. 101-108. In: Depyrogenation, Technical Report No. 7, 1985. Parenteral Drug Association, Inc., Philadelphia, PA.
10. Transfusion and Infusion Assemblies and Similar Medical Devices, chapter <161>, USP, current revision, United States Pharmacopeial Convention, Rockville, MD.

Our experienced staff will be pleased to discuss the practical and theoretical aspects of the LAL test. Please call if you have problems using Pyrotell®-T.

To download a copy of the Pyrotell®-T Instructions for Use in various languages, please visit www.accusia.com